

产 品 说 明 书

User's Guide

型号 / MODEL: CT-3086

[中文](#)


[English](#)

温馨提示

- 1.敬请在使用此仪表前，详细阅读此说明书
- 2.使用前请先清洁干净探头并在清水中浸泡5~10分钟使电极活化。
- 3.仪表开机时会显示产品型号，随后进入测量模式

[中文](#)

[English](#)



序号	描 述
1	电池区
2	数据显示区
3	温度显示区
4~6	按键区
7	固定电极环
8	电极部分

特点

➤ 测量范围 0.0 ~ 5.0 % / 0~50 PPT.

➤ LCD 液晶双显示测量值和温度值

➤ 内置电极一体式数字笔式仪表，方便使用

➤ 显示锁定功能 (HOLD)

➤ 全防水设计，体积小，重量轻

➤ °C/°F温度单位可选

➤ Nacl 4.0% 一点校准

➤ 1.5 V 电池 (LR44, BAT) x 4 PCs

➤ 电极可拆卸，更换简单方便

➤ 自动温度补偿，在温度变化的情况下保证精度

➤ 出厂前已校正，可直接使用

➤ 5分钟自动关机功能

➤ 电极内部不能有气泡，否则会引起测量误差!

➤ 产品广泛应用于环保，食品，卫生，地质探矿，冶金，地面水，工业废水，饮用水，酸雨等检测

技术参数

➤ 显示屏	➤ LCD, 尺寸 : 20 mm x 27 mm
➤ 测量范围	➤ 0.0 ~ 5.0 % / 0~50 PPT.
➤ 分辨率	➤ 0.1 % / 1 PPT.
➤ 精 度	➤ ± 2 %F·S
➤ 校 正	➤ Nacl溶液 4.0% 一点校准
➤ 工作温度	➤ 0 ~ 50°C(32 ~ 122 °F)
➤ 精 度	➤ 1°C / 2°F
➤ 自动补偿温度	➤ 0 ~ 50°C
➤ 尺寸	➤ 188 x 38mm
➤ 重量	➤ 82 g (包括电极).

操作指南

使用前请用力拨下仪表下方的电极保护帽，请勿旋动!

一、恢复出厂设置

- 1.按下“ON/OFF”键开机（如已开机，可省略）
- 2.长按“ME”键，直到显示屏显示**CLR**符号后，才可放起按键（确认仪表不能在锁定状态下）
- 3.仪表再次校正后可正常使用。

二、自动温度补偿（ATC）

在显示屏左下角有**ATC**符号出现，表示仪表处于自动温度补偿工作模式中。

三、锁定功能

按“ON/OFF键”开机后待仪表进入测量模式，按下“ME键”锁定液晶显示屏上的当前数据并且在显示屏上会出现闪烁的“hold”符号，再次点按“ME 键”即解除锁定模式。

四、仪表校正（首先准备校正时用标准缓冲液-以4.0% 为例，温度在25°C为佳）

- 1.校正点为： 4.0%；请准备正确的缓冲液；
- 2.仪表在校正过程中，未结束前切勿将仪表从校正液中取出
 - a.显示屏-左下角“**CAL**”符号显示中，此刻仪表处于校正模式


b.显示屏还没有显示“END”符号前，此时仪表还处于校正过程中

操作步骤（为保证仪表测量精度请即时进行仪表校正操作，如果仪表长时间未使用，请先将仪表电极部分浸泡在蒸馏水或纯净水中10~30分钟，使电极活化，活化过程中仪表可处于关机模式）


步 骤	描 述
1	电极活化后，点按“ON/OFF键”开机
2	将电极插入标准缓冲液中，适当搅动后静止等待读数稳定
3	长按“CAL键”大约三秒，“CAL”符号出现在显示屏上
4	此时放起“CAL键”，等待仪表自动识别当前标准缓冲液的值并显示
5	之后显示“SA”符号，仪表校正成功并存储正确数据
6	1秒后显示符号“END”，仪表退出校正模式
提示	如果仪表跳过第5步，直接执行第6步，表示仪表校正失败，需重新校正或恢复出厂设置后再校正仪表。

五、功能切换

功能切换/Function switch




型号/MODEL:CT-3086



型号/MODEL:CT-3088

温度切换/Temperature switch (“Hold/ME”键选择, “CAL/FUN”键确认)



六、超量程显示

当测量值低于 0 或者高于 5.5% 时，测量值显示区会显示：“1 - . -”。

当温度低于0°C或高于50°C时，温度显示区会显示：符号 ‘L’ 或符号 ‘H’ 。

七、低电压提示

当电池符号显示或显示屏闪烁时，表示电池电压已不足，请尽快更换新电池，切记新旧电池不要混用。

八、产品保修

产品购买之日起，免费保修一年（不包含电池和玻璃电极）

在下列情况下不属于保修范围：

使用不当造成损坏（如电池漏液、摔坏、进水等）；工作温度超过50°C，外观受到损坏、超过保修时间以及产品被自行拆装；

九、温馨提示

1. 仪表属于专业仪表，非专业人士请在专业人士指导下进行操作，以免造成仪器故障！
2. 首次使用前请把电极用清水浸泡10分钟左右，
3. 仪表标定时请用标准缓冲液标定以及正确的操作方法，不正确的标定将会引起测量误差超标！在非标定状态下，请勿按“CAL”键，否则会引起仪器错误的标定，可能引起仪器不能正常工作或测量误差严重超标！
4. 使用时请不要把电导率笔插入液体过深，只要液体能够没过电极即可。使用前特别注意电极固定环是否松动，如果电极固定环松动，电导率笔使用过程中测试的液体可能会进入仪器内部造成故障。

Tips:


1. Please read this manual carefully before using this instrument

Before use, please clean the probe thoroughly and soak it in clean water for 5-10 minutes to activate the electrode.

When the instrument is turned on, it will display the product model and then enter the measurement mode

[中文](#)

[English](#)



No,	Description
1	Battery Area
2	Data Display Area
3	Temperature Area
4~6	Key Area
7	Fixed electrode ring
8	Electrode Part

Specification

➤ Measurement range 0.0 ~ 5.0 % / 0~50 PPT	➤ LCD dual display measurement and temperature values
➤ Built in electrode integrated digital pen instrument, convenient to use	➤ Display lock function (HOLD)
➤ Fully waterproof design, small size, light weight	➤ Temperature units of °C/°F are optional
➤ NaCl 4.0% single point calibration	➤ 1.5 V (LR44, BAT) x 4 PCs
➤ The electrode is detachable and easy to replace	➤ Automatic temperature compensation ensures accuracy under temperature changes
➤ It has been calibrated before leaving the factory and can be used directly	➤ 5 minute automatic shutdown function
➤ There should be no bubbles inside the electrode, otherwise it will cause measurement errors!	
➤ The product is widely used in environmental protection, food, hygiene, geological exploration, metallurgy, surface water, industrial wastewater, drinking water, acid rain and other testing	

Features

➤ LCD	➤ LCD, Size : 20 mm x 27 mm
➤ Measurement Range	➤ 0.0 ~ 5.0 % / 0~50 PPT.
➤ Resolution	➤ 0.1 % / 1 PPT.
➤ Accuracy	➤ $\pm 2 \%F \cdot S$
➤ Calibration	➤ NaCl solution 4.0% point calibration
➤ Working Temperature	➤ 0 ~ 50°C(32 ~ 122 °F)
➤ Resolution	➤ 1°C / 2°F
➤ ATC	➤ 0 ~ 50°C
➤ Size	➤ 188 x 38mm
➤ Weight	➤ 82 g (包括电极).

Operational Guidelines

Before use, please forcefully remove the electrode protection cap under the instrument panel and do not rotate it!

A、Restore factory settings

1. Press the "ON/OFF" button to turn on the device (if it is already turned on, it can be omitted)
2. Long press the "ME" button until the **CLR** symbol is displayed on the screen, then release the button (confirm that the instrument cannot be locked)
3. After recalibration, the instrument can be used normally.

B、Automatic Temperature Compensation (ATC)

The ATC symbol appears in the lower left corner of the display screen, indicating that the instrument is in automatic temperature compensation mode.

C、Hold Function

Press the "ON/OFF" button to turn on the device and wait for the instrument to enter measurement mode.

Press the "ME" button to lock the current data on the LCD screen and a flashing "hold" symbol will appear on the screen. Press the "ME" button again to unlock the mode.

D、Instrument calibration (first prepare calibration with standard buffer solution - taking 4.0% as an example, the temperature is best at 25 °C)

1. The calibration points are: 4.0% ; Please prepare the correct buffer solution;
2. Do not remove the instrument from the calibration solution before the calibration process is completed
 - a. Display screen - The "CAL" symbol in the bottom left corner is flashing, indicating that the instrument is currently in calibration mode


b. Before the display screen shows the "END" symbol, the instrument is still in the calibration process

Operation steps (To ensure the measurement accuracy of the instrument, please perform instrument calibration immediately. If the instrument has not been used for a long time, please soak the electrode part of the instrument in distilled water or purified water for 10-30 minutes to activate the electrode. During the activation process, the instrument can be in shutdown mode)


Step	Description
1	After electrode activation, press the "ON/OFF" button to turn on the device
2	Insert the electrode into the standard buffer solution, stir it appropriately, and wait for the reading to stabilize
3	Press and hold the 'CAL' button for approximately three seconds, and the 'CAL' symbol will appear on the display screen
4	At this point, press the 'CAL' button and wait for the instrument to automatically recognize the current value of the standard buffer and display it
5	Afterwards, the "SR" symbol will be displayed, indicating that the instrument calibration has been successful and the correct data has been stored
6	After 1 second, the symbol "END" will be displayed, and the instrument will exit calibration mode
Remark	If the instrument skips step 5 and directly executes step 6, it indicates that the instrument calibration has failed and needs to be recalibrated or restored to factory settings before calibrating the instrument.

E、Function Switch

功能切换/**Function switch**




型号/MODEL:CT-3086



型号/MODEL:CT-3088

温度切换/**Temperature switch** (“Hold/ME”键选择, “CAL/FUN”键确认)



F、Over range display

When the measured value is below 0 or above 5.5%, the measurement display area will show: "1-.-".

When the temperature is below 0 °C or above 50 °C, the temperature display area will show the symbol 'L' or symbol 'H'.

G、Low Voltage Indication

When the battery symbol displays or the display screen flashes, it indicates that the battery voltage is insufficient. Please replace the new battery as soon as possible and remember not to mix old and new batteries.

H、Product Warranty

Free one-year warranty from the date of purchase (excluding battery and glass electrode)

The following situations are not covered by the warranty:

Damage caused by improper use (such as battery leakage, breakage, water ingress, etc.); The working temperature exceeds 50 °C, the appearance is damaged, the warranty period has expired, and the product has been disassembled by oneself;

I、Remind

1. Instruments belong to professional instruments. Non professionals are advised to operate under the guidance of professionals to avoid instrument malfunctions!

2. Before the first use, please soak the electrode in clean water for about 10 minutes,

3. When calibrating the instrument, please use standard buffer solution and correct operating methods. Incorrect calibration will cause measurement errors to exceed the standard! Do not press the "CAL" button in non calibrated state, otherwise it may cause incorrect calibration of the instrument, which may result in the instrument not working properly or measurement errors exceeding the standard seriously!

4. When using, please do not insert the conductivity pen too deeply into the liquid, as long as the liquid can pass through the electrode. Pay special attention to whether the electrode fixing ring is loose before use. If the electrode fixing ring is loose, the liquid tested during the use of the conductivity pen may enter the instrument and cause malfunction.